

JULIO 2021

NUEVA PRUEBA INTERLABORATORIO FAPAS

El Laboratorio de Química Analítica de Agidea (**AgideaLabs**), participó por segunda vez de la prueba internacional de aptitud interlaboratorio, llevada a cabo por la prestigiosa organización FAPAS de Inglaterra.

Los 70 participantes de esta edición recibieron muestras de harina de avena para analizar 4 analitos: Chlormequat, Mepiquat, Glifosato y AMPA. FAPAS evalúa y califica los resultados de los distintos laboratorios otorgando un puntaje (Z-Score) y desvío respecto a los valores de referencia, e informando el ranking de los puntajes de manera codificada para mantener la confidencialidad. Estas pruebas de aptitud son consideradas por las organizaciones de acreditación como evidencia externa de aptitud.

Nuevamente el resultado fue satisfactorio en cada uno de los 4 analitos:

Analito	Z-Score
Mepiquat	0,6
Chlormequat	1,2
Glifosato	-0,3
AMPA	-0,1
Z-Score Satisfactorio $ \leq 2 $	



CERTIFICATE OF PARTICIPATION

This certificate confirms that:

AGIDEA S.R.L

took part in:

Food Chemistry Proficiency Test **09138**

Start Date: 28/04/2021

Chlormequat (sum of chlormequat and its salts; expressed as chlormequat chloride) , Mepiquat (sum of mepiquat and its salts; expressed as mepiquat chloride) , Glyphosate , AMPA in Oats

and were allocated laboratory number **31**.

CEERTIFICADO DE PARTICIPACIÓN EN LA PRUEBA DE APTITUD INTERLABORATORIO

Matriz: Avena

Analitos: Chlormequat, Mepiquat, Glifosato y AMPA